

TSS 500N4

TELECOM SURGE GENERATOR



FOR TESTS ACCORDING TO ...

- › EN 60950-1
- › EN 61000-4-5
- › FCC 97-270 (part 68)
- › IEC 60065
- › IEC 60950-1
- › IEC 61000-4-5
- › ITU-T K.17
- › ITU-T K.20
- › ITU-T K.28
- › ITU-T K.45

TSS 500N4 - COMPACT TELECOM SURGE GENERATOR WITH 10/700US TEST PULSE







Telecommunication networks are exposed to lightning events. Therefore telecommunications equipment being connected to the outside world need to have appropriate protection to show an acceptable immunity to surge transients in order not to fail in case of lightning events. Telecom Surge simulators of the TSS 500N-series are used to proof the immunity of telecommunications equipment.

The TSS 500N4 is used to perform tests as per IEC 61000-4-5 and related standards and complies with the requirements of ITU-T and FCC 97-270 (part 68) for Surge B pulse.

HIGHLIGHTS

- › Standalone tester for 10/700µs pulse as per IEC 61000-4-5
- › Complies to ITU-T
- › Complies to FCC part 68 (Surge B pulses)
- › Built-in CDNs for 2-wire and 4-wire applications
- › Manual operation
- › Standard Test routines

APPLICATION AREAS

- | | |
|--|---|
|  INDUSTRY |  TELECOM |
|  COMPONENTS |  RESIDENTIAL |
|  MEDICAL | |
|  BROADCAST | |

TECHNICAL DETAILS

AS PER ITU AND ETS RECOMMENDATIONS	
	Pulse 1.2/50µs
Voltage (o.c.)	160V - 4,000V ±10%
Rise time*)	1.0µs ± 30%
Pulse duration*)	50µs ± 20%
Energy storage capacitor	1µF
	Pulse 10/700µs
Rise time*)	6.5µs ± 30%
Pulse duration*)	700µs ± 20%
Energy storage capacitor	20µF
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless
	*) definition of waveform parameters as per IEC 60469-1. As per IEC 61000-4-5 this is considered to be equal to the waveform parameter definition as per IEC 60060-1 for the 1.2/50µs pulse and CCITT for the 10/700µs pulse.

AS PER FCC PART 68, PULSE B	
Voltage (o.c.)	160V - 4,000V ±10%
Front time	9µs ± 30%
Decay time	720µs ± 20%
Current (s.c.)	4A - 100A
Front time	5µs ± 30%
Decay time	320µs ± 20%
Energy storage capacitor	20µF
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless

AS PER IEC 61000-4-5	
	Pulse 10/700µs
Voltage (o.c.)	160V - 4,000V ±10%
Rise time*)	6.5µs ± 30%
Pulse duration*)	700µs ± 20%
Current (s.c.)	4A - 100A
Rise time*)	4µs ± 20%
Pulse duration*)	300µs ± 20%
Energy storage capacitor	20µF
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless
	*) definition of waveform parameters as per IEC 60469-1. As per IEC 61000-4-5 this is considered to be equal to the waveform parameter definition as per IEC 60060-1 for the 1.2/50µs pulse and CCITT for the 10/700µs pulse.

COUPLING AS PER	
ITU-T	2-wire: T1 and T2 with 25ohm each
FCC part 68	2 wire: T1 and T2 with 25ohm each
IEC 61000-4-5	4-wire: T1,T2,T3,T4 with 100ohm each

TRIGGER	
Automatic	Automatic pulse release
Manual	Single pulse release
External	External pulse release
CRO trigger	5V trigger signal for oscilloscope

TEST ROUTINES	
Quick Start	Immediate start; easy to use and fast
User Test routines	Change Polarity after n pulses Change voltage after n pulses
Standard Test routines	IEC 61000-4-5 Level 1,000V IEC 61000-4-5 Level 2,000V IEC 61000-4-5 Level 4,000V FCC part 68, Pulse B Metallic 1,000V FCC part 68, Pulse B Longitudinal 1,500V
Service	Service, setup, self test

TECHNICAL DETAILS

INTERFACE

Serial interface	USB
Parallel interface	IEEE 488, addresses 1 - 30

SAFETY

Safety circuit	Control input (24Vdc)
Warning lamp	Floating output contact

GENERAL DATA

Dimensions, weight	19"/3HU, approx. 20kg
Supply voltage	115/230V +10/-15%
Fuses	2xT 2AT (230V) or 2xT4AT (115V)

OPTIONS

CNV 504S1	4 telecom lines as per fig. 14, IEC 61000-4-5
CNV 508S1	8 telecom lines as per fig. 14, IEC 61000-4-5
iec.control	Software to control the test, including standard library, test report facility and data conversion generator

COMPETENCE WHEREVER YOU ARE



CONTACT EM TEST DIRECTLY

Switzerland

EM TEST (Switzerland) GmbH › Sternenhofstraße 15 › 4153 Reinach › Switzerland
 Phone +41 (0)61/7179191 › Fax +41 (0)61/7179199
 Internet: www.emtest.ch › E-mail: sales.emtest@ametek.com

Germany

EM TEST GmbH › Lünener Straße 211 › 59174 Kamen › Deutschland
 Phone +49 (0)2307/26070-0 › Fax +49 (0)2307/17050
 Internet: www.emtest.com › E-mail: info.emtest@ametek.de

France

EM TEST FRANCE › Le Trident - Parc des Collines › Immeuble B1 - Etage 3 › 36, rue Paul Cézanne › 68200 Mulhouse › France
 Phone +33 (0)389 31 23 50 › Fax +33 (0)389 31 23 55
 Internet: www.emtest.fr › E-mail: info@emtest.fr

Poland

EM TEST Polska › ul. Ogrodowa 31/35, 00-893 Warszawa › Polska
 Phone +48 (0)518 64 35 12
 Internet: www.emtest.com/pl › E-mail: info_polska.emtest@ametek.de

USA / Canada

EM TEST USA › 9250 Brown Deer Road › San Diego › CA 92121
 Phone +1 (858) 699 1685 › Fax +1 (858) 458 0267
 Internet: www.emtest.com › E-mail: sales.emtest@ametek.com

P.R. China

E & S Test Technology Limited › Rm 913, Leftbank › No. 68 Bei Si Huan Xi Lu › Haidian District › Beijing 100080 › P.R. China
 Phone +86 (0)10 82 67 60 27 › Fax +86 (0)10 82 67 62 38
 Internet: www.emtest.com › E-mail: info@emtest.com.cn

Republic of Korea

EM TEST Korea Limited › #405 › WooYeon Plaza › #986-8 › YoungDeok-dong › Giheung-gu › Yongin-si › Gyeonggi-do › Korea
 Phone +82 (31) 216 8616 › Fax +82 (31) 216 8616
 Internet: www.emtest.co.kr › E-mail: sales@emtest.co.kr

Information about scope of delivery, visual design and technical data correspond with the state of development at time of release. \nTechnical data subject to change without further notice.